

Runyes Autoclave Certificate of Validation



Registered business	Job address	Landlord/agent
Excel Health Care Limited Unit 5 Childerditch Industrial Park Childerditch Hall Drive Little Warley Brentwood Essex CM13 3HD +44 0208 597 1404	MARCH HOUSE 2 SUSSEX STREET BOGNOR REGIS WEST SUSSEX PO21 1SF	N/A
Question		
Certificate of validation, a series of tests are carried out as instructed under HTM 01-05 guidelines. This certificate is valid for 1 year only from the date of testing.		
PASSED		
Autoclave Type & Model		
Enigma 17LN		
Serial Number		
1101025		
Max W.P.		
2.7 BAR		
Automatic Air Leakage Test / Certified fit for use (E13)		
N/A		
Residual Air Test / Certified fit for use (E5)		
Passed		
Automatic Control Test / Certified fit for use (E4)		
Passed		
Steam Penetration Test / Advised customer in the use of		
Failed		
Thermometric Test / Certified fit for use (E1/E2/E3)		
Passed		
Air Leakage Test / Advised customer if fitted		
N/A		
Steam Generator cut out test / Certified fit for use		



Runyes Autoclave Certificate of Validation (9890)

Registered business :Excel Health Care Limited

Registered business address :Unit 5 Childerditch Industrial Park Childerditch Hall Drive Little Warley Brentwood Essex CM13 3HD

Runyes Autoclave Certificate of Validation



Question		
Passed		
Test Date		
26 Jul 2022		
Engineer Name		
S Coultham		
Engineer information	Received by information	Certificate information
Engineer name: Steven Coultham (PO4) (M,W,F) Signature: 	Name: DOVE HOLISTIC DENTAL CENTRE Received by: Customer Signature: 	Date: 26/07/2022 Number : 9890